

CASE NARRATIVE

Parsons Transportation Group of New York

Project Name: Zerega Bus Depot

Project # N/A

Chemtech Project # J3790

Test Name: PESTICIDE Group1

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 06/28/2018.

2 Water samples were received on 06/28/2018.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Cyanide, Flash Point, Herbicide, Ignitability, Mercury, Metals Group1, Metals ICP-Group1, PCB, PESTICIDE Group1, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOCMS Group3, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TPH GC and VOCMS Group3. This data package contains results for PESTICIDE Group1.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of PESTICIDE Group1s was based on method 8081B and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples

The RPD recoveries met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____