

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071320\
Data File : PD058595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2020 14:06
Operator : AJ\MA
Sample : L3249-02MS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BFXP7MS

Manual Integrations
APPROVED

Ankita
7/14/2020 11:32:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 04:29:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(17) 4,4'-DDT (MA)
6.037min 103.599 ng/ml
response 147033631

(17) 4,4'-DDT #2 (MA)
6.962min 88.581 ng/ml
response 199708669

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071320\
Data File : PD058595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2020 14:06
Operator : AJ\MA
Sample : L3249-02MS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BFXP7MS

Manual Integrations
APPROVED

Ankita
7/14/2020 11:32:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 04:29:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(17) 4,4'-DDT (MA)

6.037min 103.599 ng/ml

response 147033631

(17) 4,4'-DDT #2 (MA)

6.960min 91.923 ng/ml m

response 207244503

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071320\
 Data File : PD058595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2020 14:06
 Operator : AJ\MA
 Sample : L3249-02MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BFXP7MS

Manual Integrations
 APPROVED

Ankita
 7/14/2020 11:32:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 04:29:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 02 16:12:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.210	3.866	32660408	50467708	19.431	22.147
27) SA Decachlor...	7.848	8.863	38143757	44560682	21.576	20.974
Target Compounds						
3) MA gamma-BHC...	3.904	4.528	102.0E6	97375677	43.745	31.417 #
4) MA Heptachlor	4.183	5.030	84551577	135.7E6	53.411	46.676
5) MB Aldrin	4.418	5.329	95804492	130.7E6	44.845	47.098
13) MA Dieldrin	5.415	6.321	210.3E6	272.2E6	121.959	99.968
14) MA Endrin	5.666	6.537	171.7E6	230.0E6	116.214	100.715
17) MA 4,4'-DDT	6.037	6.960	147.0E6	207.2E6	103.599	91.923m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

3
 AJ
 07/16/20