

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071420\
 Data File : PD058608.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2020 13:26
 Operator : AJ\MA
 Sample : L3260-09
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 X2M79

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 01:00:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 02 16:12:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.213	3.865	27345731	57341755	16.269	25.164 #
2)	SA Decachlor...	7.852	8.863	50013348	58112343	28.290	27.352
Target Compounds							
2)	A alpha-BHC	3.636	4.249	155.8E6	207.3E6	62.631	61.260
3)	MA gamma-BHC...	3.908	4.530	112.5E6	150.7E6	48.234	48.606
4)	MA Heptachlor	4.187	5.030	73831000	114.1E6	46.638	39.262
5)	MB Aldrin	4.422	5.330	52992696	70744617	24.805	25.492
6)	B beta-BHC	4.157	4.701	37806276	57823310	39.891	41.825
7)	B delta-BHC	4.348	4.914	79604857	82862505	34.575	37.406
8)	B Heptachlo...	4.855	5.714	60939950	83327354	34.553	32.215
9)	A Endosulfan I	5.183	6.067	103.8E6	134.9E6	57.725	55.253
10)	B trans-Chl...	5.075	5.947	78988270	104.1E6	40.097	40.223
11)	B cis-Chlor...	5.132	6.018	106.4E6	139.5E6	53.722	53.173
12)	B 4,4'-DDE	5.301	6.179	143.1E6	183.3E6	71.209	71.408
13)	MA Dieldrin	5.421	6.322	95595585	131.5E6	55.433	48.280
14)	MA Endrin	5.676	6.538	210.1E6	148.6E6	142.212	65.066 #
16)	A 4,4'-DDD	5.809	6.666	73148870	110.2E6	60.527	50.723
17)	MA 4,4'-DDT	6.043	6.962	163.6E6	231.0E6	115.293	102.470
19)	B Endosulfa...	6.323	7.093	45014281	64607127	29.659	28.656

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071420\
Data File : PD058608.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2020 13:26
Operator : AJ\MA
Sample : L3260-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 01:00:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
X2M79