

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071422\
 Data File : PD070867.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 18:39
 Operator : AR\AJ
 Sample : RESC033
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESC035

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 07:26:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 14 05:14:58 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.474	4.125	48398566	467.3E6	20.115	19.913
27)	SA Decachlor...	8.245	9.322	104.4E6	397.8E6	41.109	43.069
Target Compounds							
2)	A alpha-BHC	3.915	4.530	34838456	309.3E6	9.287	10.248
3)	MA gamma-BHC...	4.199	4.823	36287521	266.9E6	9.968	10.313
4)	MA Heptachlor	4.499	5.340	32503038	219.7E6	9.328	10.960
5)	MB Aldrin	4.748	5.652	37338546	200.4E6	10.333	11.514
6)	B beta-BHC	4.449	5.000	16683654	143.6E6	10.649	10.739
7)	B delta-BHC	4.654	5.224	35524066	256.8E6	10.197	11.321
8)	B Heptachlo...	5.193	6.047	34434990	165.0E6	10.342	11.113
9)	A Endosulfan I	5.532	6.412	28935228	134.3E6	9.599	10.131
10)	B trans-Chl...	5.419	6.284	35681078	164.6E6	10.252	10.931
11)	B cis-Chlor...	5.479	6.357	34872095	155.3E6	10.204	10.752
12)	B 4,4'-DDE	5.645	6.512	65463766	299.4E6	20.346	21.830
13)	MA Dieldrin	5.777	6.673	63084839	283.2E6	18.838	20.451
14)	MA Endrin	6.035	6.896	53528792	204.8E6	19.309	19.618
15)	B Endosulfa...	6.309	7.110	57094749	258.1E6	20.819	22.410
16)	A 4,4'-DDD	6.162	7.012	49851995	229.5E6	19.142	20.987
17)	MA 4,4'-DDT	6.404	7.313	50188339	234.2E6	18.758	20.928
18)	B Endrin al...	6.479	7.236	49255986	223.1E6	21.502	22.943
19)	B Endosulfa...	6.692	7.460	58862293	256.0E6	21.013	22.577
20)	A Methoxychlor	6.955	7.772	139.0E6	562.2E6	100.529	107.456
21)	B Endrin ke...	7.186	7.942	68043653	284.6E6	21.950	23.048

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071422\
Data File : PD070867.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 18:39
Operator : AR\AJ
Sample : RESC033
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 07:26:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

