



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

EA Engineering Science & Technology

Project Name: Scotia, NY - Annual Testing

Project # N/A

Chemtech Project # O3532

Test Name: PCB

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 07/06/2023.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: PCB, TCLP Extraction and TCLP PCB. This data package contains results for PCB.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_Q. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 μ m df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 μ m; Catalogue # 7HM-G017-11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

Sample FSND-IDW-S fax was sent with NON-DOD criteria and sample has no any positive detection of PCB. However this sample needs to be reported with DOD criteria. Lab has reported correct data with DOD criteria in hard copy.

The not QT review data is reported in the Miscellaneous.

The soil samples results are based on a dry weight basis.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____