

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071423\
 Data File : PD076739.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2023 06:32
 Operator : AR\AJ
 Sample : 03444-17
 Misc :
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 A46X8

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/17/2023
 Supervised By : Ankita Jodhani 07/17/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 09:51:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 14:58:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.765	2474.7E6	9873823	1163.259	9.250 #
27)	SA Decachlor...	9.083	7.904	57764573	22136464	23.776	20.650
Target Compounds							
3)	MA gamma-BHC...	4.338	3.574	1863.9E6	1598.3E6	561.930m	1066.848m#
7)	B delta-BHC	4.801	4.131	1881.3E6	179.8E6	561.008m	117.805 #
8)	B Heptachlo...	5.686	4.710	729.2E6	403.9E6	243.389m	287.097m
10)	B trans-Chl...	5.971	4.945	273.9E6	244.3E6	89.076m	180.634m#
11)	B cis-Chlor...	6.032	5.017	153.1E6	66399146	48.288	47.064
14)	MA Endrin	6.585	5.631	413.4E6	51213478	160.356	41.415m#
17)	MA 4,4'-DDT	7.006	6.017	390.8E6	175.3E6	175.016m	187.125m
19)	B Endosulfa...	7.150	6.319	46457128	37839914	18.603m	33.915m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071423\
Data File : PD076739.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2023 06:32
Operator : AR\AJ
Sample : 03444-17
Misc :
ALS Vial : 77 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 09:51:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 14:58:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 07/17/2023
Supervised By: Ankita Jodhani 07/17/2023

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
A46X8