

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071425\
 Data File : PD089477.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2025 14:11
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/16/2025
 Supervised By :mohammad ahmed 07/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 07:23:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 04:35:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.557	2.880	146.8E6	897.0E6	51.304	51.321
7) SA Decachlor...	9.081	8.073	179.6E6	947.5E6	45.906	45.870
Target Compounds						
2) Toxaphene-1	6.249	5.474	18005870	80844564	539.550	494.704
3) Toxaphene-2	6.450	5.646	26097738	55036726	568.801	509.872
4) Toxaphene-3	7.156	6.757	42465876	239.0E6	492.674	479.937
5) Toxaphene-4	7.571	7.198	51532642	165.3E6	465.284	475.877
6) Toxaphene-5	7.935	7.329	30703702	104.5E6	494.456m	430.154

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071425\
Data File : PD089477.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2025 14:11
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 07:23:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 04:35:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 07/16/2025
Supervised By: mohammad ahmed 07/22/2025

Instrument :
ECD_D
Client Sampled :
PTOXCCC500