

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Method\
 Method File : PD071618.M
 Title : GC Extractables
 Last Update : Tue Jul 17 02:39:14 2018
 Response Via : Initial Calibration

Calibration Files

50 =PD048396.D 100 =PD048394.D 75 =PD048395.D
 25 =PD048397.D 5 =PD048398.D

	Compound	50	100	75	25	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.339	1.379	1.353	1.310	1.399	1.356	E6	2.54
2) MA	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	gamma-Chlordane						0.000		-1.00
11) B	alpha-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Mirex						0.000		-1.00
23)	Chlordane-1	5.712	5.807	5.733	5.708	6.663	5.925	E4	7.00
24)	Chlordane-2	5.805	5.791	5.763	5.785	6.373	5.904	E4	4.45
25)	Chlordane-3	1.936	1.969	1.945	1.821	2.001	1.935	E5	3.52
26)	Chlordane-4	1.664	1.697	1.663	1.653	1.713	1.678	E5	1.52
27)	Chlordane-5	5.896	5.938	5.894	5.842	5.934	5.901	E4	0.66
28) SA	Decachlorobiphenyl	1.457	1.450	1.447	1.462	1.594	1.482	E6	4.22

Signal #2 Calibration Files

50 =PD048396.D 100 =PD048394.D 75 =PD048395.D
 25 =PD048397.D 5 =PD048398.D

	Compound	50	100	75	25	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.902	1.856	1.871	1.936	2.148	1.943	E7	6.12
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	gamma-Chlordane						0.000		-1.00
11) B	alpha-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Method\
 Method File : PD071618.M
 Title : GC Extractables
 Last Update : Tue Jul 17 02:39:14 2018
 Response Via : Initial Calibration

Calibration Files

50 =PD048396.D 100 =PD048394.D 75 =PD048395.D
 25 =PD048397.D 5 =PD048398.D

	Compound	50	100	75	25	5	Avg	%RSD
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00
23)	Chlordane-1	0.958	0.969	0.983	1.037	1.061	1.002 E6	4.50
24)	Chlordane-2	0.940	0.909	0.912	0.954	1.034	0.950 E6	5.36
25)	Chlordane-3	3.338	3.251	3.291	3.415	3.970	3.453 E6	8.55
26)	Chlordane-4	3.845	3.729	3.789	3.998	4.716	4.015 E6	10.07
27)	Chlordane-5	9.207	8.566	8.806	8.826	9.841	9.049 E5	5.51
28) SA	Decachlorobiphenyl	1.901	1.805	1.841	1.934	2.209	1.938 E7	8.25

(#) = Out of Range