

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Method\
 Method File : DTX071618.M
 Title : GC Extractables
 Last Update : Tue Jul 17 03:53:29 2018
 Response Via : Initial Calibration

Calibration Files

500 =PD048401.D 1000 =PD048399.D 750 =PD048400.D
 250 =PD048402.D 50 =PD048403.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.273	1.285	1.263	1.240	1.266	1.265	E6	1.31
2)	Toxaphene-1	0.950	1.017	0.947	0.952	1.043	0.982	E4	4.58
3)	Toxaphene-2	2.673	2.648	2.617	2.631	2.969	2.707	E4	5.45
4)	Toxaphene-3	2.764	2.779	2.783	2.784	2.956	2.813	E4	2.85
5)	Toxaphene-4	2.255	2.271	2.299	2.231	2.048	2.221	E4	4.48
6)	Toxaphene-5	3.157	3.175	3.137	2.923	2.177	2.914	E4	14.56
7) SA	Decachlorobiphenyl	1.455	1.436	1.423	1.447	1.531	1.459	E6	2.90

Signal #2 Calibration Files

500 =PD048401.D 1000 =PD048399.D 750 =PD048400.D
 250 =PD048402.D 50 =PD048403.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.911	1.870	1.866	1.925	1.957	1.906	E7	2.02
2)	Toxaphene-1	1.045	1.070	1.156	1.253	0.950	1.095	E5	10.50
3)	Toxaphene-2	2.086	2.060	2.034	1.935	2.188	2.061	E5	4.44
4)	Toxaphene-3	5.815	5.471	5.463	5.632	6.140	5.704	E5	4.96
5)	Toxaphene-4	3.511	3.329	3.379	3.567	4.078	3.573	E5	8.36
6)	Toxaphene-5	2.355	2.223	2.239	2.185	2.377	2.276	E5	3.74
7) SA	Decachlorobiphenyl	1.899	1.809	1.818	1.923	2.178	1.925	E7	7.76

(#) = Out of Range