

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071723\
 Data File : PD076817.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jul 2023 11:38
 Operator : AR\AJ
 Sample : 03444-19DL 40X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 13:47:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071423CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 14 14:58:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	2.764	61429946	396514	28.876	0.371 #
27)	SA Decachlor...	9.029f	7.907	694350	1381181	0.286	1.288 #
Target Compounds							
2)	A alpha-BHC	3.950f	0.000	296.8E6	0	83.724	N.D. #
3)	MA gamma-BHC...	4.339	3.574	73277802	-1080779	22.091	N.D. #
4)	MA Heptachlor	0.000	3.950	0	192.6E6	N.D.	128.827
5)	MB Aldrin	5.295	4.231	8157823	42923868	2.386	28.521 #
6)	B beta-BHC	4.511f	3.850f	52744363	51169160	34.701	71.427 #
7)	B delta-BHC	4.818f	4.130	90470175	6483641	26.978	4.248 #
8)	B Heptachlo...	5.687	4.711	29759396	8484412	9.932	6.031 #
9)	A Endosulfan I	6.066	5.058	3904299	2834147	1.371	2.287 #
10)	B trans-Chl...	5.973	4.946	-5352004	2922106	N.D.	2.161
11)	B cis-Chlor...	6.032	5.058	5078722	2834147	1.602	2.009 #
12)	B 4,4'-DDE	6.198	5.208	8921946	322322	3.062	0.267 #
13)	MA Dieldrin	6.355	5.353	4139822	14166314	1.312	10.159 #
14)	MA Endrin	6.586	5.633	22384937	-6141686	8.684	N.D. #
15)	B Endosulfa...	6.855f	5.910	7239484	3523200	2.723	2.929
16)	A 4,4'-DDD	6.732	5.837f	9145657	3292044	4.047	3.474
17)	MA 4,4'-DDT	7.097f	6.019	4622180	7351038	2.070	7.847 #
18)	B Endrin al...	0.000	6.119	0	109175	N.D.	0.130
19)	B Endosulfa...	7.203	6.320	1242841	1482929	0.498	1.329 #
20)	A Methoxychlor	7.447f	6.631	1482873	258892	1.120	0.485 #
22)	Toxaphene-1	6.247	4.995	1070046	13333700	48.165	1812.286 #
24)	Toxaphene-3	6.664	5.968	2091646	1406746	108.020	153.422 #
25)	Toxaphene-4	7.153	6.631f	1426763	258892	23.791	10.319 #
26)	Toxaphene-5	0.000	6.709	0	262329	N.D.	7.085

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071723\
Data File : PD076817.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jul 2023 11:38
Operator : AR\AJ
Sample : 03444-19DL 40X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 13:47:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071423CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jul 14 14:58:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
0