

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM140


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024

Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) alpha-BHC (A)

4.018min 9.689 ng/ml

response 18329653

(2) alpha-BHC #2 (A)

3.271min 10.362 ng/ml

response 93412640

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM140


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024

Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) alpha-BHC (A)

4.018min 9.689 ng/ml

response 18329653

(2) alpha-BHC #2 (A)

3.270min 10.153 ng/ml m

response 91525329

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
PEM140

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024
Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.611min 46.674 ng/ml

response 77196473

(14) Endrin #2 (MA)

5.626min 45.305 ng/ml

response 300036144

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
PEM140

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024
Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.609min 47.852 ng/ml m

response 79144841

(14) Endrin #2 (MA)

5.626min 45.305 ng/ml

response 300036144

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
PEM140

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024
Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

6.744min 1.084 ng/ml

response 1564084

(16) 4,4'-DDD #2 (A)

0.000min 0.000 ng/ml

response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071724\
Data File : PD083988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2024 17:17
Operator : AR\AJ
Sample : PEM140
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_D
LabSampleId :
PEM140

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/17/2024
Supervised By :Ankita Jodhani 07/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:04:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD062624CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

6.744min 1.084 ng/ml

response 1564084

(16) 4,4'-DDD #2 (A)

5.771min 1.002 ng/ml m

response 5969847