

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071919\
 Data File : PD054015.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jul 2019 1:07
 Operator : SM\AJ
 Sample : K3591-02
 Misc : CHLOR LOQ 50 PPB
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 LOQ-SOIL-02-QT3-2019

Manual Integrations
 APPROVED

Ankita
 7/19/2019 12:04:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 06:39:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071919.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 04:33:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	3.982	12037169	19041640	20.915	21.859
28)	SA Decachlor...	8.007	9.024	14714690	21317639	21.859	22.995
Target Compounds							
23)	Chlordane-1	4.159	4.972	1082448	2101261	54.706	56.089
24)	Chlordane-2	4.651	5.436	1598545	2569957	61.761m	60.636m
25)	Chlordane-3	5.210	6.080	4352575	9640023	54.728	55.049
26)	Chlordane-4	5.266	6.154	3989349	11892803	57.965	56.184
27)	Chlordane-5	6.090	6.955	1227841	2216986	45.969m	58.969 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071919\
Data File : PD054015.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2019 1:07
Operator : SM\AJ
Sample : K3591-02
Misc : CHLOR LOQ 50 PPB
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
LOQ-SOIL-02-QT3-2019

Manual Integrations
APPROVED

Ankita
7/19/2019 12:04:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 06:39:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071919.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 04:33:23 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

