

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071922\
 Data File : PD070999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2022 23:39
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD071922

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 05:05:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071922.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 20 05:04:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.474	4.125	143.8E6	1396.0E6	49.609	48.283
28)	SA Decachlor...	8.242	9.321	149.8E6	555.4E6	49.042	48.698
Target Compounds							
2)	A alpha-BHC	3.916	4.530	221.5E6	1820.0E6	50.476	48.017
3)	MA gamma-BHC...	4.200	4.824	206.0E6	1443.8E6	47.454	48.342
4)	MA Heptachlor	4.500	5.340	209.4E6	1058.6E6	50.296	47.921
5)	MB Aldrin	4.748	5.653	210.6E6	882.4E6	50.393	47.665
6)	B beta-BHC	4.449	5.000	82144142	598.3E6	49.147	52.485
7)	B delta-BHC	4.654	5.224	201.1E6	1134.6E6	50.813	47.810
8)	B Heptachlo...	5.193	6.047	187.5E6	730.9E6	50.046	47.826
9)	A Endosulfan I	5.533	6.412	175.1E6	696.5E6	50.158	48.002
10)	B gamma-Chl...	5.419	6.284	194.0E6	754.5E6	49.904	47.953
11)	B alpha-Chl...	5.478	6.357	190.5E6	734.7E6	49.817	48.047
12)	B 4,4'-DDE	5.644	6.512	177.9E6	725.2E6	50.503	48.315
13)	MA Dieldrin	5.777	6.674	189.8E6	738.3E6	50.517	48.320
14)	MA Endrin	6.035	6.896	157.9E6	541.7E6	50.562	52.228
15)	B Endosulfa...	6.308	7.111	149.5E6	600.3E6	49.711	47.859
16)	A 4,4'-DDD	6.161	7.012	147.6E6	610.8E6	50.296	49.103
17)	MA 4,4'-DDT	6.402	7.313	155.2E6	652.0E6	50.691	48.680
18)	B Endrin al...	6.478	7.236	117.7E6	488.0E6	49.081	48.388
19)	B Endosulfa...	6.691	7.460	148.1E6	588.2E6	49.706	48.678
20)	A Methoxychlor	6.953	7.773	81352795	333.6E6	49.211	48.549
21)	B Endrin ke...	7.185	7.942	173.7E6	664.8E6	50.147	48.036
22)	Mirex	7.371	8.396	134.0E6	444.4E6	48.323	48.358

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071922\
Data File : PD070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2022 23:39
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD071922

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 05:05:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071922.M
Quant Title : GC Extractables
QLast Update : Wed Jul 20 05:04:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

