

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071922\
 Data File : PD071030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2022 08:07
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/20/2022
 Supervised By : Ankita Jodhani 07/20/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 08:40:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071922.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 20 05:30:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.474	4.125	149.2E6	1361.5E6	51.471	47.089
28)	SA Decachlor...	8.242	9.322	149.7E6	554.7E6	49.001	48.639
Target Compounds							
2)	A alpha-BHC	3.916	4.530	229.6E6	1776.6E6	52.327	46.872
3)	MA gamma-BHC...	4.199	4.824	205.2E6	1423.3E6	47.249	47.655
4)	MA Heptachlor	4.498	5.340	206.2E6	1050.5E6	49.519m	47.555
5)	MB Aldrin	4.747	5.653	209.3E6	876.5E6	50.086m	47.348
6)	B beta-BHC	4.449	5.000	81179987	586.4E6	48.570	43.719
7)	B delta-BHC	4.654	5.224	202.2E6	1143.7E6	51.096	48.372
8)	B Heptachlo...	5.192	6.047	185.0E6	707.8E6	49.377	46.812
9)	A Endosulfan I	5.530	6.412	171.6E6	689.7E6	49.152m	47.513
10)	B gamma-Chl...	5.419	6.284	192.6E6	751.0E6	49.539	47.615
11)	B alpha-Chl...	5.478	6.357	189.7E6	728.9E6	49.619	47.374
12)	B 4,4'-DDE	5.644	6.512	180.7E6	715.0E6	51.324	47.633
13)	MA Dieldrin	5.776	6.673	193.5E6	735.7E6	51.515	48.145
14)	MA Endrin	6.035	6.896	157.9E6	541.9E6	50.579	52.247
15)	B Endosulfa...	6.308	7.110	148.7E6	598.7E6	49.421	47.827
16)	A 4,4'-DDD	6.161	7.012	148.2E6	612.9E6	50.488	49.278
17)	MA 4,4'-DDT	6.403	7.313	150.8E6	640.9E6	49.264	47.849
18)	B Endrin al...	6.478	7.236	115.9E6	486.4E6	48.327	48.235
19)	B Endosulfa...	6.690	7.460	147.2E6	588.0E6	49.416m	48.660
20)	A Methoxychlor	6.953	7.772	78171050	328.6E6	47.287	47.816
21)	B Endrin ke...	7.185	7.942	179.7E6	662.8E6	51.871	47.888
22)	Mirex	7.370	8.396	132.8E6	445.9E6	47.912m	48.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD071922\
Data File : PD071030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2022 08:07
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/20/2022
Supervised By :Ankita Jodhani 07/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 08:40:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071922.M
Quant Title : GC Extractables
QLast Update : Wed Jul 20 05:30:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

