

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072019\
 Data File : PD054041.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jul 2019 13:28
 Operator : SM\AJ
 Sample : K3800-09MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 A52G7MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 23:20:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071819CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 07:29:30 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.314	3.981	6211158	9224933	10.097	10.066
27)	SA Decachlor...	8.010	9.025	11458513	16374209	16.363	17.441
Target Compounds							
3)	MA gamma-BHC...	4.021	4.651	23190397	34950476	26.900	28.123
4)	MA Heptachlor	4.309	5.159	21266066	35746362	28.199	27.642
5)	MB Aldrin	4.550	5.463	22194420	35034267	26.304	26.804
13)	MA Dieldrin	5.562	6.458	49209373	74545947	55.961	56.580
14)	MA Endrin	5.816	6.675	44969635	61291775	59.920	59.343
17)	MA 4,4'-DDT	6.187	7.094	36544440	57586106	59.608	56.618

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072019\
Data File : PD054041.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2019 13:28
Operator : SM\AJ
Sample : K3800-09MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
A52G7MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 23:20:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071819CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 07:29:30 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

