

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072320\
 Data File : PD058670.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2020 14:51
 Operator : AJ\MA
 Sample : INDA309
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA309

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 06:16:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 02 16:12:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.208	3.866	28237830	41002142	16.800	17.994
27)	SA Decachlor...	7.844	8.859	58499217	69342606	33.090	32.638
Target Compounds							
2)	A alpha-BHC	3.631	4.248	42724426	60149514	17.170	17.771
3)	MA gamma-BHC...	3.902	4.528	40331651	55476015	17.296	17.899
4)	MA Heptachlor	4.180	5.029	30483226	53328203	19.256	18.346
9)	A Endosulfan I	5.175	6.064	31427321	43769696	17.477	17.926
13)	MA Dieldrin	5.413	6.319	66778328	96909193	38.723	35.592
14)	MA Endrin	5.663	6.535	55062965	80353820	37.275	35.180
16)	A 4,4'-DDD	5.801	6.664	47342246	76626803	39.173	35.259
17)	MA 4,4'-DDT	6.034	6.960	52050630	78807951	36.674	34.955
20)	A Methoxychlor	6.583	7.420	108.0E6	202.8E6	198.081	173.383

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072320\
Data File : PD058670.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2020 14:51
Operator : AJ\MA
Sample : INDA309
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 06:16:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

