

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD D\Data\PD072418\
 Data File : PD048473.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jul 2018 12:57
 Operator : AJ\SJ
 Sample : INDA350
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA350

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 04:50:18 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 29 16:27:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.393	4.007	25847989	399.7E6	21.390	19.633
27) SA Decachlor...	8.081	9.049	55232544	745.1E6	41.316	42.048
Target Compounds						
2) A alpha-BHC	3.825	4.390	41248289	656.6E6	21.680	20.427
3) MA gamma-BHC...	4.103	4.672	38889833	577.8E6	20.739	19.919
4) MA Heptachlor	4.391	5.186	38640798	594.7E6	21.840	20.557
9) A Endosulfan I	5.404	6.226	32209086	474.6E6	21.243	20.173
13) MA Dieldrin	5.644	6.481	72700613	970.0E6	42.043	40.196
14) MA Endrin	5.898	6.696	62975402	800.3E6	41.826	40.524
16) A 4,4'-DDD	6.027	6.813	57887529	782.7E6	41.825	41.701
17) MA 4,4'-DDT	6.263	7.113	59274908	783.0E6	40.945	40.840
20) A Methoxychlor	6.811	7.565	151.2E6	1700.0E6	199.828	200.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD072418\
Data File : PD048473.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jul 2018 12:57
Operator : AJ\SJ
Sample : INDA350
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA350

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 04:50:18 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\ECD_D\Method\PD062818CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 29 16:27:50 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

