

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072520\
Data File : PD058679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 17:01
Operator : AJ\MA
Sample : L3388-04MS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
C0AD1MS

Manual Integrations
APPROVED

Ankita
7/27/2020 12:23:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 00:03:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(17) 4,4'-DDT (MA)

6.034min 1.014 ng/ml

response 1439071

(17) 4,4'-DDT #2 (MA)

6.961min 1.016 ng/ml

response 2290453

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072520\
Data File : PD058679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 17:01
Operator : AJ\MA
Sample : L3388-04MS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
C0AD1MS

Manual Integrations
APPROVED

Ankita
7/27/2020 12:23:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 00:03:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(17) 4,4'-DDT (MA)
6.033min 1.163 ng/ml m
response 1650252

(17) 4,4'-DDT #2 (MA)
6.961min 1.016 ng/ml
response 2290453

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072520\
Data File : PD058679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 17:01
Operator : AJ\MA
Sample : L3388-04MS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C0AD1MS

Manual Integrations
APPROVED

Ankita
7/27/2020 12:23:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 00:03:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.209min 14.811 ng/ml

response 24893563

(1) Tetrachloro-m-xylene #2 (SA)

3.866min 15.290 ng/ml

response 34840957

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072520\
Data File : PD058679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 17:01
Operator : AJ\MA
Sample : L3388-04MS
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
C0AD1MS

Manual Integrations
APPROVED

Ankita
7/27/2020 12:23:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 00:03:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 02 16:12:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.209min 14.811 ng/ml

response 24893563

(1) Tetrachloro-m-xylene #2 (SA)

3.864min 15.951 ng/ml m

response 36347685

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072520\
 Data File : PD058679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2020 17:01
 Operator : AJ\MA
 Sample : L3388-04MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AD1MS

Manual Integrations
 APPROVED

Ankita
 7/27/2020 12:23:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 00:03:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD070220CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 02 16:12:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.209	3.864	24893563	36347685	14.811	15.951m
27) SA Decachlor...	7.845	8.860	30135422	37596260	17.046	17.696
Target Compounds						
17) MA 4,4'-DDT	6.033	6.961	1650252	2290453	1.163m	1.016

AJ
 07/31/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.