

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072524\
 Data File : PD084097.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2024 18:22
 Operator : AR\AJ
 Sample : INDB201
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 03:50:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072524CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 25 03:48:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.561	2.771	9523459	57151923	10.027	10.286
27)	SA Decachlor...	9.112	7.894	40752347	127.2E6	20.531	20.909
Target Compounds							
5)	MB Aldrin	5.287	4.214	18623845	81863336	10.005	10.121
6)	B beta-BHC	4.549	3.895	7863062	38402680	10.196	10.300
7)	B delta-BHC	4.798	4.124	19591245	87627729	10.449	10.045
8)	B Heptachlo...	5.715	4.715	16937212	76651086	9.831	10.244
10)	B trans-Chl...	5.972	4.964	18106443	74238943	10.079	10.167
11)	B cis-Chlor...	6.051	5.028	18613176	76047708	10.127	10.239
12)	B 4,4'-DDE	6.223	5.216	36902023	143.4E6	20.028	20.284
15)	B Endosulfa...	6.828	5.917	35544046	127.0E6	20.512	20.623
18)	B Endrin al...	6.959	6.096	25793678	94102522	20.189	20.508
19)	B Endosulfa...	7.194	6.319	33299703	121.2E6	20.252	20.482
21)	B Endrin ke...	7.680	6.825	36145872	130.8E6	19.987	20.368


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072524\
Data File : PD084097.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2024 18:22
Operator : AR\AJ
Sample : INDB201
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 03:50:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072524CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 25 03:48:48 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.561 min
Delta R.T.: 0.000 min
Response: 9523459
Conc: 10.03 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 57151923
Conc: 10.29 ng/ml




#5 Aldrin

R.T.: 5.287 min
Delta R.T.: 0.000 min
Response: 18623845
Conc: 10.00 ng/ml




#5 Aldrin

R.T.: 4.214 min
Delta R.T.: 0.000 min
Response: 81863336
Conc: 10.12 ng/ml




#6 beta-BHC

R.T.: 4.549 min
Delta R.T.: 0.000 min
Response: 7863062
Conc: 10.20 ng/ml




#6 beta-BHC

R.T.: 3.895 min
Delta R.T.: 0.000 min
Response: 38402680
Conc: 10.30 ng/ml




#7 delta-BHC

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 19591245
Conc: 10.45 ng/ml




#7 delta-BHC

R.T.: 4.124 min
Delta R.T.: 0.000 min
Response: 87627729
Conc: 10.04 ng/ml




#8 Heptachlor epoxide

R.T.: 5.715 min
Delta R.T.: 0.000 min
Response: 16937212
Conc: 9.83 ng/ml




#8 Heptachlor epoxide

R.T.: 4.715 min
Delta R.T.: 0.000 min
Response: 76651086
Conc: 10.24 ng/ml




#10 trans-Chlordane

R.T.: 5.972 min
Delta R.T.: 0.000 min
Response: 18106443
Conc: 10.08 ng/ml




#10 trans-Chlordane

R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 74238943
Conc: 10.17 ng/ml




#11 cis-Chlordane

R.T.: 6.051 min
Delta R.T.: 0.000 min
Response: 18613176
Conc: 10.13 ng/ml




#11 cis-Chlordane

R.T.: 5.028 min
Delta R.T.: 0.000 min
Response: 76047708
Conc: 10.24 ng/ml




#12 4,4'-DDE

R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 36902023
Conc: 20.03 ng/ml




#12 4,4'-DDE

R.T.: 5.216 min
Delta R.T.: 0.000 min
Response: 143355772
Conc: 20.28 ng/ml




#15 Endosulfan II

R.T.: 6.828 min
Delta R.T.: 0.000 min
Response: 35544046
Conc: 20.51 ng/ml




#15 Endosulfan II

R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 126975497
Conc: 20.62 ng/ml




#18 Endrin aldehyde

R.T.: 6.959 min
Delta R.T.: 0.000 min
Response: 25793678
Conc: 20.19 ng/ml



#18 Endrin aldehyde

R.T.: 6.096 min
Delta R.T.: 0.000 min
Response: 94102522
Conc: 20.51 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.194 min

Delta R.T.: 0.000 min

Response: 33299703

Conc: 20.25 ng/ml




#19 Endosulfan Sulfate

R.T.: 6.319 min

Delta R.T.: 0.000 min

Response: 121216776

Conc: 20.48 ng/ml




#21 Endrin ketone

R.T.: 7.680 min

Delta R.T.: 0.000 min

Response: 36145872

Conc: 19.99 ng/ml




#21 Endrin ketone

R.T.: 6.825 min

Delta R.T.: 0.000 min

Response: 130774942

Conc: 20.37 ng/ml




#27 Decachlorobiphenyl

R.T.: 9.112 min

Delta R.T.: 0.000 min

Response: 40752347

Conc: 20.53 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.894 min

Delta R.T.: 0.000 min

Response: 127193688

Conc: 20.91 ng/ml