

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072525\
 Data File : PD089624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2025 11:15
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 23:30:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 22 04:39:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.879	158.1E6	1057.5E6	54.858	54.049
28)	SA Decachlor...	9.067	8.067	204.3E6	1169.2E6	49.754	47.590
Target Compounds							
2)	A alpha-BHC	3.997	3.390	326.1E6	1637.0E6	57.122	54.486
3)	MA gamma-BHC...	4.328	3.727	309.2E6	1494.1E6	56.037	53.819
4)	MA Heptachlor	4.926	4.079	286.3E6	1445.3E6	52.493	51.585
5)	MB Aldrin	5.268	4.365	290.0E6	1466.3E6	55.428	53.543
6)	B beta-BHC	4.513	4.022	118.4E6	629.9E6	54.924	52.921
7)	B delta-BHC	4.761	4.258	296.9E6	1498.4E6	56.169	53.714
8)	B Heptachlo...	5.687	4.868	257.5E6	1327.0E6	54.523	53.375
9)	A Endosulfan I	6.071	5.242	240.2E6	1241.3E6	54.035	52.276
10)	B gamma-Chl...	5.942	5.121	261.6E6	1451.3E6	55.177	54.248
11)	B alpha-Chl...	6.023	5.186	257.3E6	1372.2E6	53.984	53.202
12)	B 4,4'-DDE	6.192	5.370	229.9E6	1384.5E6	53.553	52.732
13)	MA Dieldrin	6.344	5.508	256.8E6	1410.0E6	54.409	53.355
14)	MA Endrin	6.571	5.785	215.2E6	1293.6E6	52.842	53.496
15)	B Endosulfa...	6.783	6.076	207.2E6	1196.5E6	50.929	52.092
16)	A 4,4'-DDD	6.702	5.925	189.2E6	1174.8E6	55.840	53.589
17)	MA 4,4'-DDT	7.018	6.179	171.3E6	1062.9E6	45.399	45.410
18)	B Endrin al...	6.912	6.254	148.9E6	840.9E6	51.080	50.662
19)	B Endosulfa...	7.146	6.478	195.0E6	1123.0E6	51.669	50.592
20)	A Methoxychlor	7.490	6.749	87638467	541.6E6	43.712	44.422
21)	B Endrin ke...	7.627	6.987	206.8E6	1225.6E6	51.391	50.048
22)	Mirex	8.109	7.181	152.1E6	955.3E6	49.451	49.421

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072525\
Data File : PD089624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2025 11:15
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 23:30:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072125.M
Quant Title : GC Extractables
QLast Update : Tue Jul 22 04:39:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

