

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX072622.M
 Title : GC Extractables
 Last Update : Tue Jul 26 04:20:10 2022
 Response Via : Initial Calibration

Calibration Files

500 =PD071161.D 1000 =PD071163.D 750 =PD071162.D
 250 =PD071160.D 100 =PD071159.D

	Compound	500	1000	750	250	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.726	2.667	2.682	2.577	2.503	2.631	E6	3.42
2)	Toxaphene-1	2.066	1.883	1.875	1.795	1.936	1.911	E4	5.24
3)	Toxaphene-2	3.620	3.649	3.673	3.651	4.164	3.752	E4	6.17
4)	Toxaphene-3	5.782	5.667	5.701	5.614	6.333	5.819	E4	5.04
5)	Toxaphene-4	6.097	5.951	5.984	5.714	5.825	5.914	E4	2.50
6)	Toxaphene-5	6.352	6.278	6.326	5.775	5.196	5.986	E4	8.37
7) SA	Decachlorobiphenyl	2.620	2.514	2.546	2.538	2.624	2.568	E6	1.95

Signal #2 Calibration Files

500 =PD071161.D 1000 =PD071163.D 750 =PD071162.D
 250 =PD071160.D 100 =PD071159.D

	Compound	500	1000	750	250	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.695	2.542	2.596	2.653	2.788	2.655	E7	3.55
2)	Toxaphene-1	4.107	3.789	3.844	4.080	3.867	3.937	E4	3.70
3)	Toxaphene-2	1.261	1.193	1.202	1.368	1.380	1.281	E5	6.96
4)	Toxaphene-3	2.332	2.157	2.241	2.343	2.510	2.316	E5	5.69
5)	Toxaphene-4	1.287	1.209	1.253	1.256	1.391	1.279	E5	5.36
6)	Toxaphene-5	1.687	1.542	1.598	1.724	1.797	1.670	E5	6.06
7) SA	Decachlorobiphenyl	0.962	0.888	0.917	0.950	1.002	0.944	E7	4.62

(#) = Out of Range