

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD072718\
 Data File : PD048577.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Jul 2018 21:33
 Operator : AJ\SJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD072718TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 10:15:21 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX072718.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 30 10:14:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.371	3.969	59739968	879.0E6	49.638	50.137
7) SA Decachlor...	8.069	8.995	76718164	871.5E6	49.454	49.179
Target Compounds						
2) Toxaphene-1	5.588	6.336	4469523	47397411	526.035	491.267
3) Toxaphene-2	6.185	7.029	12324404	90265194	522.519	509.801
4) Toxaphene-3	6.429	7.104	13682181	240.4E6	478.815	509.045
5) Toxaphene-4	6.781	7.191	11895129	141.4E6	538.780	472.964
6) Toxaphene-5	6.911	7.596	16993008	97028213	494.847	476.604

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD072718\
Data File : PD048577.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Jul 2018 21:33
Operator : AJ\SJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
ICVPD072718TOX

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 10:15:21 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX072718.M
Quant Title : GC Extractables
QLast Update : Mon Jul 30 10:14:51 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

