

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD072818CLP.M
 Title : GC Extractables
 Last Update : Sat Jul 28 02:39:26 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD048596.D 1 =PD048594.D 2 =PD048595.D
 4 =PD048597.D 5 =PD048598.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.338	1.370	1.363	1.303	1.276	1.330	E6	3.02
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	1.029	1.075	1.003	1.017	0.955	1.016	E4	4.25
23)	Toxaphene-2	2.723	2.741	2.788	2.806	2.730	2.758	E4	1.34
24)	Toxaphene-3	3.035	3.013	3.016	3.014	2.993	3.014	E4	0.49
25)	Toxaphene-4	2.600	2.609	2.524	2.595	2.573	2.580	E4	1.33
26)	Toxaphene-5	3.522	3.669	3.596	3.617	3.530	3.587	E4	1.71
27) SA	Decachlorobiphenyl	1.584	1.678	1.622	1.545	1.487	1.583	E6	4.61

Signal #2 Calibration Files

3 =PD048596.D 1 =PD048594.D 2 =PD048595.D
 4 =PD048597.D 5 =PD048598.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.882	2.007	1.898	1.748	1.673	1.841	E7	7.16
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD072818CLP.M
 Title : GC Extractables
 Last Update : Sat Jul 28 02:39:26 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD048596.D 1 =PD048594.D 2 =PD048595.D
 4 =PD048597.D 5 =PD048598.D

Compound		3	1	2	4	5	Avg	%RSD	
21) B	Endrin ketone						0.000	-1.00	
22)	Toxaphene-1	0.997	1.082	1.091	0.931	0.929	1.006	E5	7.82
23)	Toxaphene-2	1.481	1.539	1.500	1.351	1.278	1.430	E5	7.72
24)	Toxaphene-3	5.058	5.386	5.240	4.733	4.510	4.985	E5	7.24
25)	Toxaphene-4	3.008	3.187	3.085	2.814	2.706	2.960	E5	6.67
26)	Toxaphene-5	1.996	2.143	2.062	1.847	1.782	1.966	E5	7.60
27) SA	Decachlorobiphenyl	1.805	1.997	1.886	1.628	1.523	1.768	E7	10.87

(#) = Out of Range