

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(11) cis-Chlordane (B)

5.474min 1.272 ng/ml

response 4348108

(11) cis-Chlordane #2 (B)

6.356min 0.762 ng/ml

response 11008541

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(11) cis-Chlordane (B)

5.472min 0.418 ng/ml m

response 1429048

(11) cis-Chlordane #2 (B)

6.356min 0.762 ng/ml

response 11008541

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)
5.640min 2.116 ng/ml
response 6809024

(12) 4,4'-DDE #2 (B)
6.511min 2.960 ng/ml
response 40594952

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)
5.639min 1.515 ng/ml m
response 4874101

(12) 4,4'-DDE #2 (B)
6.511min 2.960 ng/ml
response 40594952

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.158min 0.653 ng/ml

response 1700494

(16) 4,4'-DDD #2 (A)

7.011min 0.720 ng/ml

response 7878343

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072822\
Data File : PD071212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2022 13:51
Operator : AR\AJ
Sample : N3741-07
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
H0B15

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 07/28/2022
Supervised By :Ankita Jodhani 07/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 00:12:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071422CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 14 05:14:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(16) 4,4'-DDD (A)

6.157min 0.754 ng/ml m

response 1964032

(16) 4,4'-DDD #2 (A)

7.010min 1.275 ng/ml m

response 13943254