

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072921\
 Data File : PD064536.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 06:59
 Operator : AR\AJ
 Sample : M3082-18
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 07:46:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.413	4.619	41308677	159.4E6	34.629m	17.446 #
27)	SA Decachlor...	11.307	10.364	30986432	146.8E6	23.539m	24.693
Target Compounds							
3)	MA gamma-BHC...	6.393	5.735	1504739	27927249	0.959	2.449 #
7)	B delta-BHC	6.914	6.388	2699336	17861291	1.551m	1.644m
11)	B cis-Chlor...	8.227	7.391	16071718	20856232	10.315	2.345m#
12)	B 4,4'-DDE	8.436	7.607	4358383	28788875	3.024m	3.815m#
13)	MA Dieldrin	8.572	7.739	10335116	73044156	6.930m	9.094m#
14)	MA Endrin	8.817	8.037	3328987	34415466	2.600	5.991m#
15)	B Endosulfa...	9.030	8.333	6651161	36764496	5.122m	6.369
16)	A 4,4'-DDD	8.962	8.180	26761795	35694637	21.801m	6.197m#
17)	MA 4,4'-DDT	9.278	8.440	6212148	169.1E6	4.759m	30.715m#
19)	B Endosulfa...	9.425	8.753	10232924	181.1E6	7.425	33.604m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072921\
Data File : PD064536.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 06:59
Operator : AR\AJ
Sample : M3082-18
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 07:46:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

