

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072921\
 Data File : PD064543.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 08:50
 Operator : AR\AJ
 Sample : M3082-02
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0002

Manual Integrations
 APPROVED

mohammad
 7/30/2021 4:50:37 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 09:15:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	5.405	4.619	16667883	151.7E6	13.973	16.603
27)	SA Decachlor...	11.307	10.363	31163685	164.0E6	23.674m	27.575
Target Compounds							
3)	MA gamma-BHC...	6.391	5.734	1123894	45621961	0.717	4.001 #
10)	B trans-Chl...	8.132	7.303	34924271	20653262	21.794m	2.277m#
13)	MA Dieldrin	8.570	7.738	3242909	14897576	2.175m	1.855m
14)	MA Endrin	8.818	8.037	932743	8522197	0.728	1.484m#
15)	B Endosulfa...	9.030	8.335	2205978	9035882	1.699m	1.565
16)	A 4,4'-DDD	8.960	8.176	11153933	11845216	9.086m	2.057 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072921\
Data File : PD064543.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 08:50
Operator : AR\AJ
Sample : M3082-02
Misc :
ALS Vial : 63 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 09:15:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD D
ClientSampleId :
C0002

Manual Integrations
APPROVED

mohammad
7/30/2021 4:50:37 PM