

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073021\
 Data File : PD064552.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 11:41
 Operator : AR\AJ
 Sample : M3082-15DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0013DL

Manual Integrations
 APPROVED

mohammad
 8/2/2021 12:35:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 01:03:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	5.408	4.617	9925705	85019498	8.321	9.304
27)	SA Decachlor...	11.307	10.363	18543970	94057433	14.087m	15.816
Target Compounds							
7)	B delta-BHC	6.916	6.388	471084	13374245	0.271	1.231 #
8)	B Heptachlo...	7.878	7.037	4254134	15684429	2.478	1.694m#
10)	B trans-Chl...	8.139	7.304	34837098	54357151	21.740m	5.993m#
11)	B cis-Chlor...	8.229	7.389	3012255	6100788	1.933m	0.686m#
12)	B 4,4'-DDE	8.423	7.603	16373595	104.1E6	11.360	13.794
13)	MA Dieldrin	8.573	7.737	7617868	36297208	5.108	4.519m
14)	MA Endrin	8.815	8.040	2068277	48289496	1.615	8.406 #
15)	B Endosulfa...	9.031	8.332	4862593	21939194	3.744m	3.801
16)	A 4,4'-DDD	8.963	8.187	106.8E6	392.0E6	87.001m	68.054
17)	MA 4,4'-DDT	9.280	8.444	40255435	327.6E6	30.836m	59.510m#
19)	B Endosulfa...	9.424	8.750	11220466	66418667	8.141	12.322m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073021\
Data File : PD064552.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 11:41
Operator : AR\AJ
Sample : M3082-15DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 01:03:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



8/2/2021 12:35:48 PM

Manual Integrations
APPROVED

mohammad

Instrument :
ECD_D
ClientSampled :
C0013DL