

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073021\
 Data File : PD064631.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2021 09:49
 Operator : AR\AJ
 Sample : M3123-22
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0083

Manual Integrations
 APPROVED

mohammad
 8/2/2021 12:38:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 06:10:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	5.405	4.616	22190254	183.6E6	18.602	20.098m
27)	SA Decachlor...	11.304	10.361	39956918	183.2E6	30.353	30.811
Target Compounds							
3)	MA gamma-BHC...	6.403	5.734	1903439	150.1E6	1.214	13.165 #
7)	B delta-BHC	6.908	6.357	4986193	79179790	2.865	7.289m#
9)	A Endosulfan I	8.258	7.464	7504190	221.4E6	5.323	28.795 #
10)	B trans-Chl...	8.130	7.331	139.4E6	27461415	86.988m	3.028m#
14)	MA Endrin	8.823	8.025	13639720	19919465	10.653	3.468 #
15)	B Endosulfa...	9.039	8.336	8857640	22130849	6.821	3.834m#
16)	A 4,4'-DDD	8.949	8.175	43269285	188.9E6	35.249	32.806
19)	B Endosulfa...	9.425	8.745	5551785	20920491	4.028m	3.881m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073021\
Data File : PD064631.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2021 09:49
Operator : AR\AJ
Sample : M3123-22
Misc :
ALS Vial : 63 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 06:10:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



8/2/2021 12:38:43 PM
mohammad

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
C0083