

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073024\
Data File : PD084141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2024 09:02
Operator : AR\AJ
Sample : PB162346BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PBLK346

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 04:36:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072424CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 25 11:17:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.562	2.771	18152689	98972332	20.274	18.974
27)	SA Decachlor...	9.118	7.896	39065607	123.5E6	20.353	21.431


Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD073024\
Data File : PD084141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2024 09:02
Operator : AR\AJ
Sample : PB162346BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 04:36:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072424CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jul 25 11:17:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Instrument :
ECD_D
ClientSampled :
PBLK346



#1 Tetrachloro-m-xylene


R.T.: 3.562 min
Delta R.T.: 0.002 min
Response: 18152689
Conc: 20.27 ng/ml

Instrument :
ECD_D
ClientSampleId :
PBLK346




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 98972332
Conc: 18.97 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.118 min
Delta R.T.: 0.005 min
Response: 39065607
Conc: 20.35 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.896 min
Delta R.T.: 0.001 min
Response: 123544300
Conc: 21.43 ng/ml