

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD073118.M
 Title : GC Extractables
 Last Update : Wed Aug 01 03:05:41 2018
 Response Via : Initial Calibration

Calibration Files

50 =PD048644.D 100 =PD048644.D 75 =PD048647.D
 25 =PD048647.D 5 =PD048648.D

	Compound	50	100	75	25	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.220	1.242	1.216	1.179	1.260	1.223	E6	2.49
2) MA	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	gamma-Chlordane						0.000		-1.00
11) B	alpha-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Mirex						0.000		-1.00
23)	Chlordane-1	5.157	5.114	5.062	5.095	4.786	5.043	E4	2.93
24)	Chlordane-2	5.234	5.103	5.099	5.196	5.271	5.181	E4	1.50
25)	Chlordane-3	1.591	1.653	1.847	1.643	1.598	1.667	E5	6.29
26)	Chlordane-4	1.496	1.499	1.478	1.455	1.543	1.494	E5	2.16
27)	Chlordane-5	5.664	5.684	5.560	5.268	5.274	5.490	E4	3.74
28) SA	Decachlorobiphenyl	1.422	1.389	1.399	1.407	1.497	1.423	E6	3.06

Signal #2 Calibration Files

50 =PD048646.D 100 =PD048644.D 75 =PD048645.D
 25 =PD048647.D 5 =PD048648.D

	Compound	50	100	75	25	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.366	1.333	1.341	1.362	1.463	1.373	E7	3.81
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	gamma-Chlordane						0.000		-1.00
11) B	alpha-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD073118.M
 Title : GC Extractables
 Last Update : Wed Aug 01 03:05:41 2018
 Response Via : Initial Calibration

Calibration Files

50 =PD048644.D 100 =PD048644.D 75 =PD048647.D
 25 =PD048647.D 5 =PD048648.D

	Compound	50	100	75	25	5	Avg	%RSD
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00
23)	Chlordane-1	7.151	6.954	6.958	6.993	8.659	7.343 E5	10.08
24)	Chlordane-2	7.236	6.725	6.827	7.280	6.316	6.877 E5	5.78
25)	Chlordane-3	2.392	2.325	2.331	2.409	2.727	2.437 E6	6.83
26)	Chlordane-4	2.752	2.652	2.664	2.780	3.281	2.826 E6	9.22
27)	Chlordane-5	6.178	6.018	6.018	5.954	6.346	6.103 E5	2.61
28) SA	Decachlorobiphenyl	1.485	1.403	1.421	1.527	1.756	1.518 E7	9.36

(#) = Out of Range