

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Method\
 Method File : DTX073118.M
 Title : GC Extractables
 Last Update : Wed Aug 01 10:17:31 2018
 Response Via : Initial Calibration

Calibration Files

500 =PD048651.D 1000 =PD048649.D 750 =PD048650.D
 250 =PD048652.D 50 =PD048653.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.147	1.150	1.148	1.123	1.205	1.155	E6	2.64
2)	Toxaphene-1	2.377	2.437	2.422	2.295	1.715	2.249	E4	13.50
3)	Toxaphene-2	4.870	4.971	4.978	4.786	4.863	4.894	E4	1.66
4)	Toxaphene-3	8.414	8.449	8.577	8.206	6.795	8.088	E4	9.09
5)	Toxaphene-4	4.608	4.528	4.566	4.536	4.375	4.523	E4	1.95
6)	Toxaphene-5	4.077	4.080	4.124	4.054	3.406	3.948	E4	7.71
7) SA	Decachlorobiphenyl	1.416	1.383	1.409	1.431	1.530	1.434	E6	3.95

Signal #2 Calibration Files

500 =PD048651.D 1000 =PD048649.D 750 =PD048650.D
 250 =PD048652.D 50 =PD048653.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.410	1.350	1.378	1.429	1.490	1.411	E7	3.78
2)	Toxaphene-1	4.111	3.980	4.089	3.979	4.374	4.106	E5	3.93
3)	Toxaphene-2	4.559	4.348	4.520	4.705	4.968	4.620	E5	5.03
4)	Toxaphene-3	3.168	3.059	3.108	3.159	3.616	3.222	E5	6.97
5)	Toxaphene-4	4.516	4.413	4.535	4.252	5.667	4.677	E5	12.08
6)	Toxaphene-5	4.204	4.076	4.169	4.298	4.586	4.267	E5	4.58
7) SA	Decachlorobiphenyl	1.584	1.491	1.529	1.663	1.937	1.641	E7	10.82

(#) = Out of Range