

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD073120CLP.M
 Title : GC Extractables
 Last Update : Thu Jul 30 02:04:06 2020
 Response Via : Initial Calibration

Calibration Files

3 =PD058793.D 1 =PD058791.D 2 =PD058792.D
 4 =PD058794.D 5 =PD058795.D

Compound			3	1	2	4	5	Avg		%RSD
1)	SA	Tetrachloro-m-xylene	1.582	1.525	1.598	1.542	1.538	1.557	E6	2.03
2)	A	alpha-BHC						0.000		-1.00
3)	MA	gamma-BHC (Lindane)						0.000		-1.00
4)	MA	Heptachlor						0.000		-1.00
5)	MB	Aldrin						0.000		-1.00
6)	B	beta-BHC						0.000		-1.00
7)	B	delta-BHC						0.000		-1.00
8)	B	Heptachlor epoxide						0.000		-1.00
9)	A	Endosulfan I						0.000		-1.00
10)	B	trans-Chlordane						0.000		-1.00
11)	B	cis-Chlordane						0.000		-1.00
12)	B	4,4'-DDE						0.000		-1.00
13)	MA	Dieldrin						0.000		-1.00
14)	MA	Endrin						0.000		-1.00
15)	B	Endosulfan II						0.000		-1.00
16)	A	4,4'-DDD						0.000		-1.00
17)	MA	4,4'-DDT						0.000		-1.00
18)	B	Endrin aldehyde						0.000		-1.00
19)	B	Endosulfan Sulfate						0.000		-1.00
20)	A	Methoxychlor						0.000		-1.00
21)	B	Endrin ketone						0.000		-1.00
22)		Toxaphene-1	1.025	0.942	1.019	1.016	1.035	1.007	E4	3.69
23)		Toxaphene-2	1.130	1.057	1.106	1.140	1.135	1.114	E4	3.06
24)		Toxaphene-3	9.243	8.574	8.846	8.701	8.639	8.800	E3	3.03
25)		Toxaphene-4	3.002	2.742	2.898	3.060	3.086	2.958	E4	4.75
26)		Toxaphene-5	3.738	3.388	3.600	3.878	3.939	3.709	E4	5.98
27)	SA	Decachlorobiphenyl	1.504	1.525	1.558	1.453	1.420	1.492	E6	3.71

Signal #2 Calibration Files

3 =PD058793.D 1 =PD058791.D 2 =PD058792.D
 4 =PD058794.D 5 =PD058795.D

Compound			3	1	2	4	5	Avg		%RSD
1)	SA	Tetrachloro-m-xylene	2.298	2.098	2.327	2.219	2.196	2.228	E6	4.07
2)	A	alpha-BHC						0.000		-1.00
3)	MA	gamma-BHC (Lindane)						0.000		-1.00
4)	MA	Heptachlor						0.000		-1.00
5)	MB	Aldrin						0.000		-1.00
6)	B	beta-BHC						0.000		-1.00
7)	B	delta-BHC						0.000		-1.00
8)	B	Heptachlor epoxide						0.000		-1.00
9)	A	Endosulfan I						0.000		-1.00
10)	B	trans-Chlordane						0.000		-1.00
11)	B	cis-Chlordane						0.000		-1.00
12)	B	4,4'-DDE						0.000		-1.00
13)	MA	Dieldrin						0.000		-1.00
14)	MA	Endrin						0.000		-1.00
15)	B	Endosulfan II						0.000		-1.00
16)	A	4,4'-DDD						0.000		-1.00
17)	MA	4,4'-DDT						0.000		-1.00
18)	B	Endrin aldehyde						0.000		-1.00
19)	B	Endosulfan Sulfate						0.000		-1.00
20)	A	Methoxychlor						0.000		-1.00
21)	B	Endrin ketone						0.000		-1.00
22)		Toxaphene-1	1.272	1.272	1.320	1.260	1.296	1.284	E4	1.87
23)		Toxaphene-2	1.625	1.468	1.696	1.642	1.657	1.618	E4	5.42

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD073120CLP.M
 Title : GC Extractables
 Last Update : Thu Jul 30 02:04:06 2020
 Response Via : Initial Calibration

Calibration Files

3 =PD058793.D 1 =PD058791.D 2 =PD058792.D
 4 =PD058794.D 5 =PD058795.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	2.104	1.745	2.041	1.884	1.923	1.939	E4	7.23
25)	Toxaphene-4	5.922	5.790	5.967	5.913	5.833	5.885	E4	1.22
26)	Toxaphene-5	2.644	2.566	2.230	2.591	2.476	2.501	E4	6.53
27) SA	Decachlorobiphenyl	1.908	2.191	2.051	1.787	1.694	1.926	E6	10.35

(#) = Out of Range