

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD073120CLP.M
 Title : GC Extractables
 Last Update : Thu Jul 30 03:02:49 2020
 Response Via : Initial Calibration

Calibration Files

3 =PD058801.D 1 =PD058797.D 2 =PD058799.D
 4 =PD058803.D 5 =PD058805.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.578	1.606	1.637	1.594	1.609	1.605 E6	1.33
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	2.010	2.038	2.061	2.082	2.148	2.068 E6	2.52
6) B	beta-BHC	9.422	9.883	9.869	9.305	9.260	9.548 E5	3.20
7) B	delta-BHC	2.254	2.178	2.254	2.347	2.428	2.292 E6	4.22
8) B	Heptachlor epoxide	1.791	1.809	1.849	1.837	1.888	1.835 E6	2.04
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	1.841	1.861	1.882	1.897	1.951	1.887 E6	2.22
11) B	cis-Chlordane	1.836	1.876	1.899	1.874	1.916	1.880 E6	1.61
12) B	4,4'-DDE	1.856	1.731	1.838	1.929	1.989	1.869 E6	5.23
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	1.644	1.646	1.664	1.671	1.685	1.662 E6	1.03
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.357	1.369	1.401	1.370	1.372	1.374 E6	1.20
19) B	Endosulfan Sulfate	1.627	1.630	1.661	1.644	1.655	1.643 E6	0.91
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	1.791	1.811	1.851	1.794	1.817	1.813 E6	1.33
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.566	1.717	1.693	1.539	1.523	1.608 E6	5.63

Signal #2 Calibration Files

3 =PD058801.D 1 =PD058797.D 2 =PD058799.D
 4 =PD058803.D 5 =PD058805.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.273	2.278	2.363	2.297	2.299	2.302 E6	1.56
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	2.787	2.720	2.810	2.867	2.930	2.823 E6	2.83
6) B	beta-BHC	1.336	1.366	1.394	1.320	1.316	1.347 E6	2.45
7) B	delta-BHC	2.243	2.016	2.416	2.638	2.801	2.423 E6	12.85
8) B	Heptachlor epoxide	2.512	2.615	2.660	2.511	2.510	2.562 E6	2.78
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	2.526	2.560	2.615	2.581	2.632	2.583 E6	1.66
11) B	cis-Chlordane	2.532	2.585	2.634	2.560	2.594	2.581 E6	1.49
12) B	4,4'-DDE	2.472	2.364	2.476	2.560	2.609	2.496 E6	3.77
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	2.167	2.287	2.259	2.183	2.171	2.214 E6	2.51
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.851	1.927	1.930	1.839	1.823	1.874 E6	2.71
19) B	Endosulfan Sulfate	2.183	2.230	2.255	2.186	2.171	2.205 E6	1.63
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	2.256	2.216	2.304	2.119	2.169	2.213 E6	3.27
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD073120CLP.M
Title : GC Extractables
Last Update : Thu Jul 30 03:02:49 2020
Response Via : Initial Calibration

Calibration Files

3 =PD058801.D 1 =PD058797.D 2 =PD058799.D
4 =PD058803.D 5 =PD058805.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	2.024	2.294	2.187	1.911	1.827	2.049 E6	9.39

(#) = Out of Range