

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080122\
 Data File : PD071280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2022 20:09
 Operator : AR\AJ
 Sample : N3961-05
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RVE-172

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 08/02/2022
 Supervised By :mohammad ahmed 08/02/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:10:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 26 03:39:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.815	3.574	32890053	380.8E6	11.598	13.529
28)	SA Decachlor...	8.017	9.105	34834349	135.4E6	12.560	13.082
Target Compounds							
12)	B 4,4'-DDE	5.306	6.233	7323574	47545930	2.140m	3.326 #
17)	MA 4,4'-DDT	6.119	7.069	5003466	19828541	1.695	1.520

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080122\
Data File : PD071280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2022 20:09
Operator : AR\AJ
Sample : N3961-05
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleID :
RVE-172

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 08/02/2022
Supervised By :mohammad ahmed 08/02/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:10:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
Quant Title : GC Extractables
QLast Update : Tue Jul 26 03:39:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

