

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD080221\  
 Data File : PD064666.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 02 Aug 2021 15:38  
 Operator : AR\AJ  
 Sample : INDA441  
 Misc :  
 ALS Vial : 49 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :

Manual Integrations  
 APPROVED

mohammad  
 8/3/2021 3:12:15 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 03 02:53:38 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD072721CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Jul 28 07:07:14 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm


|                             | Compound        | RT#1   | RT#2   | Resp#1   | Resp#2  | ng/ml   | ng/ml   |
|-----------------------------|-----------------|--------|--------|----------|---------|---------|---------|
| -----                       |                 |        |        |          |         |         |         |
| System Monitoring Compounds |                 |        |        |          |         |         |         |
| 1)                          | SA Tetrachlo... | 5.405  | 4.618  | 23582872 | 215.8E6 | 19.770  | 23.613  |
| 27)                         | SA Decachlor... | 11.302 | 10.363 | 43459719 | 228.5E6 | 33.014  | 38.431  |
| Target Compounds            |                 |        |        |          |         |         |         |
| 2)                          | A alpha-BHC     | 5.987  | 5.320  | 32544870 | 289.3E6 | 19.977  | 23.472  |
| 3)                          | MA gamma-BHC... | 6.383  | 5.740  | 30821022 | 256.0E6 | 19.652m | 22.445  |
| 4)                          | MA Heptachlor   | 7.034  | 6.149  | 30013706 | 243.6E6 | 17.859  | 21.481  |
| 9)                          | A Endosulfan I  | 8.276  | 7.448  | 25929025 | 157.7E6 | 18.394m | 20.508  |
| 13)                         | MA Dieldrin     | 8.568  | 7.733  | 55880718 | 309.7E6 | 37.472  | 38.553  |
| 14)                         | MA Endrin       | 8.805  | 8.023  | 42515228 | 228.2E6 | 33.205m | 39.730  |
| 16)                         | A 4,4'-DDD      | 8.960  | 8.186  | 47142186 | 236.5E6 | 38.404  | 41.058  |
| 17)                         | MA 4,4'-DDT     | 9.278  | 8.446  | 41740207 | 199.7E6 | 31.974  | 36.274  |
| 20)                         | A Methoxychlor  | 9.770  | 9.039  | 118.2E6  | 468.6E6 | 165.224 | 183.006 |
| -----                       |                 |        |        |          |         |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD080221\  
Data File : PD064666.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 02 Aug 2021 15:38  
Operator : AR\AJ  
Sample : INDA441  
Misc :  
ALS Vial : 49 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 03 02:53:38 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD072721CLP.M  
Quant Title : GC Extractables  
QLast Update : Wed Jul 28 07:07:14 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



8/3/2021 3:12:15 PM  
mohammad

Manual Integrations  
APPROVED

Instrument :  
ECD\_D  
ClientSampled :  
mohammad