

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080323\
 Data File : PD076999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2023 14:52
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICPD080323

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 17:53:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 14:17:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.762	116.4E6	66829184	49.166	49.063
28)	SA Decachlor...	9.083	7.904	137.5E6	70286262	50.180	49.114
Target Compounds							
2)	A alpha-BHC	4.000	3.264	196.7E6	98166921	51.217	50.749
3)	MA gamma-BHC...	4.333	3.594	187.7E6	94308320	50.985	50.694
4)	MA Heptachlor	4.924	3.933	184.1E6	91840127	51.058	50.554
5)	MB Aldrin	5.267	4.213	189.8E6	93587935	51.348	50.544
6)	B beta-BHC	4.530	3.890	73858169	37494051	49.081	49.146
7)	B delta-BHC	4.779	4.120	178.5E6	90743557	51.179	51.001
8)	B Heptachlo...	5.695	4.716	166.2E6	83917066	50.204	50.182
9)	A Endosulfan I	6.082	5.086	154.0E6	77730929	50.659	49.829
10)	B gamma-Chl...	5.952	4.966	164.1E6	83456107	50.428	50.311
11)	B alpha-Chl...	6.031	5.030	166.2E6	83539726	50.671	50.025
12)	B 4,4'-DDE	6.203	5.219	152.3E6	74027462	51.329	51.047
13)	MA Dieldrin	6.358	5.351	167.5E6	85241144	51.170	50.601
14)	MA Endrin	6.588	5.626	140.8E6	74443071	50.659	50.876
15)	B Endosulfa...	6.807	5.920	146.9E6	71949511	46.166	48.811
16)	A 4,4'-DDD	6.721	5.773	120.1E6	57654311	51.064	51.844
17)	MA 4,4'-DDT	7.037	6.024	127.0E6	60663998	52.229	54.574
18)	B Endrin al...	6.938	6.099	107.7E6	55911644	49.304	49.114
19)	B Endosulfa...	7.173	6.322	131.1E6	66970490	50.326	50.214
20)	A Methoxychlor	7.513	6.599	72793013	32159832	51.376	53.785
21)	B Endrin ke...	7.659	6.829	153.3E6	80894599	50.888	51.022
22)	Mirex	8.134	7.010	114.4E6	64304688	50.136	49.664

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080323\
Data File : PD076999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2023 14:52
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD080323

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 17:53:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 14:17:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

