

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080421\
 Data File : PD064728.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2021 15:14
 Operator : AR\AJ
 Sample : M3169-20
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 07:13:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.406	4.617	14910172	146.4E6	12.499	16.020m#
27)	SA Decachlor...	11.314	10.359	48222685	127.3E6	36.633	21.410 #
Target Compounds							
7)	B delta-BHC	6.908	6.359	2712033	33506374	1.558	3.084 #
8)	B Heptachlo...	7.876	7.043	4292196	11934570	2.500	1.289 #
12)	B 4,4'-DDE	8.438	7.606	4582363	40030213	3.179	5.305 #
13)	MA Dieldrin	8.569	7.738	5196980	82740867	3.485m	10.302 #
14)	MA Endrin	8.818	8.034	5231896	17234963	4.086	3.000 #
15)	B Endosulfa...	9.031	8.335	5337941	45487690	4.111	7.880 #
16)	A 4,4'-DDD	8.955	8.176	35232309	66123673	28.701m	11.481 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080421\
Data File : PD064728.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2021 15:14
Operator : AR\AJ
Sample : M3169-20
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 07:13:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

