

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080422\
 Data File : PD071325.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2022 14:06
 Operator : AR\AJ
 Sample : N3980-01
 Misc : CHLOR LOD 25 PPB
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 LOD-MDL-SOIL-01-QT3-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 03:20:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 26 03:39:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.815	3.573	46394324	521.5E6	16.359	18.526
28)	SA Decachlor...	8.016	9.103	48640053	198.7E6	17.537	19.189
Target Compounds							
23)	Chlordane-1	3.829	4.737	2128698	19657213	19.826	25.977 #
24)	Chlordane-2	4.412	5.270	2585669	15858174	22.060m	26.474
25)	Chlordane-3	5.051	5.981	7210202	55229445	21.116	29.791m#
26)	Chlordane-4	5.115	6.060	6847524	63674875	21.584	29.214m#
27)	Chlordane-5	6.019	6.921	2534954	11320113	20.966	25.028m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080422\
Data File : PD071325.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2022 14:06
Operator : AR\AJ
Sample : N3980-01
Misc : CHLOR LOD 25 PPB
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
LOD-MDL-SOIL-01-QT3-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 03:20:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
Quant Title : GC Extractables
QLast Update : Tue Jul 26 03:39:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

