

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080425\
 Data File : PD089733.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2025 16:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 06:35:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD073125.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 01 07:03:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.878	127.6E6	852.6E6	51.749	53.830
2)	SA Decachlor...	9.069	8.067	192.8E6	1112.9E6	51.877	55.431
Target Compounds							
2)	A alpha-BHC	3.997	3.390	265.3E6	1327.8E6	53.552	54.630
3)	MA gamma-BHC...	4.328	3.726	250.9E6	1221.8E6	52.802	54.243
4)	MA Heptachlor	4.926	4.079	249.6E6	1239.8E6	51.867	53.420
5)	MB Aldrin	5.268	4.365	239.1E6	1202.9E6	52.372	54.062
6)	B beta-BHC	4.514	4.022	95482987	510.8E6	51.761	52.938
7)	B delta-BHC	4.762	4.258	243.7E6	1229.0E6	53.143	54.106
8)	B Heptachlo...	5.688	4.869	215.5E6	1096.8E6	51.539	53.374
9)	A Endosulfan I	6.072	5.242	203.1E6	1038.2E6	51.242	53.108
10)	B gamma-Chl...	5.943	5.121	216.4E6	1167.7E6	51.785	53.925
11)	B alpha-Chl...	6.024	5.186	215.8E6	1119.8E6	51.325	53.780
12)	B 4,4'-DDE	6.193	5.370	194.0E6	1141.2E6	52.003	53.988
13)	MA Dieldrin	6.344	5.508	217.9E6	1169.2E6	51.591	54.032
14)	MA Endrin	6.571	5.785	179.5E6	1034.4E6	50.173	51.916
15)	B Endosulfa...	6.784	6.076	185.5E6	1015.2E6	50.551	54.000
16)	A 4,4'-DDD	6.703	5.925	155.4E6	947.4E6	51.881	53.733
17)	MA 4,4'-DDT	7.018	6.179	170.0E6	1017.9E6	50.579	55.404
18)	B Endrin al...	6.912	6.254	130.2E6	716.9E6	52.177	55.490
19)	B Endosulfa...	7.147	6.478	174.2E6	996.7E6	50.212	53.686
20)	A Methoxychlor	7.491	6.749	88908266	525.4E6	49.726	53.636
21)	B Endrin ke...	7.628	6.987	189.6E6	1112.7E6	51.052	55.445
22)	Mirex	8.110	7.180	138.8E6	842.4E6	49.784	55.346

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080425\
Data File : PD089733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2025 16:10
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 06:35:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD073125.M
Quant Title : GC Extractables
QLast Update : Fri Aug 01 07:03:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

