

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080519\
 Data File : PD054179.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Aug 2019 21:03
 Operator : SM\AJ
 Sample : K4143-07
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PST-006-B261-073019

Manual Integrations
 APPROVED

Ankita
 8/6/2019 2:26:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 04:18:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071919.M
 Quant Title : GC Extractables
 QLast Update : Fri Jul 19 04:33:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.312	3.983	12443224	18390546	21.620	21.111
28) SA Decachlor...	8.007	9.025	11175916	15217511	16.602	16.415
Target Compounds						
17) MA 4,4'-DDT	6.184	7.092	239129	597085	0.415m	0.606m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080519\
Data File : PD054179.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Aug 2019 21:03
Operator : SM\AJ
Sample : K4143-07
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PST-006-B261-073019


Manual Integrations
APPROVED

Ankita
8/6/2019 2:26:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 04:18:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD071919.M
Quant Title : GC Extractables
QLast Update : Fri Jul 19 04:33:23 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.312 min
Delta R.T.: 0.000 min
Response: 12443224
Conc: 21.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
PST-006-B261-073019


Manual Integrations
APPROVED

Ankita
8/6/2019 2:26:23 PM




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 18390546
Conc: 21.11 ng/ml




#17 4,4'-DDT

R.T.: 6.184 min
Delta R.T.: 0.000 min
Response: 239129
Conc: 0.42 ng/ml m



#17 4,4'-DDT

R.T.: 7.092 min
Delta R.T.: -0.003 min
Response: 597085
Conc: 0.61 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.007 min
Delta R.T.: -0.001 min
Response: 11175916
Conc: 16.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PST-006-B261-073019

Manual Integrations
APPROVED

Ankita
8/6/2019 2:26:23 PM



#28 Decachlorobiphenyl

R.T.: 9.025 min
Delta R.T.: 0.000 min
Response: 15217511
Conc: 16.41 ng/ml