

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080522\
 Data File : PD071367.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 15:36
 Operator : AR\AJ
 Sample : N3980-09
 Misc : CHLOR MDL 25 PPB
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 MDL-WATER-03-QT3-2022

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2022
 Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 00:45:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 26 03:39:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.814	3.573	44713245	497.0E6	15.767	17.657
28)	SA Decachlor...	8.014	9.102	47956667	200.0E6	17.291	19.316
Target Compounds							
23)	Chlordane-1	3.829	4.736	2018809	16986432	18.803	22.447
24)	Chlordane-2	4.411	5.269	2572953	16590054	21.952m	27.696 #
25)	Chlordane-3	5.050	5.980	6908743	53621393	20.233	28.923m#
26)	Chlordane-4	5.114	6.060	6664127	61780065	21.006	28.344 #
27)	Chlordane-5	6.017	6.921	2519084	9580877	20.835	21.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080522\
Data File : PD071367.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 15:36
Operator : AR\AJ
Sample : N3980-09
Misc : CHLOR MDL 25 PPB
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
MDL-WATER-03-QT3-2022

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2022
Supervised By :Ankita Jodhani 08/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 00:45:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
Quant Title : GC Extractables
QLast Update : Tue Jul 26 03:39:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

