

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080522\
 Data File : PD071370.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 16:16
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 09:36:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 26 03:39:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.814	3.573	124.3E6	1300.9E6	43.815	46.214
28)	SA Decachlor...	8.014	9.102	135.4E6	544.1E6	48.802	52.557
Target Compounds							
2)	A alpha-BHC	3.324	4.039	180.7E6	1655.1E6	43.088	44.583
3)	MA gamma-BHC...	3.658	4.375	173.5E6	1322.7E6	43.385	44.175
4)	MA Heptachlor	4.005	4.953	176.7E6	974.4E6	41.664	46.084
5)	MB Aldrin	4.289	5.296	176.2E6	819.2E6	43.202	46.388
6)	B beta-BHC	3.956	4.581	69274175	527.2E6	40.551	44.508
7)	B delta-BHC	4.190	4.829	168.8E6	1054.8E6	43.001	53.004
8)	B Heptachlo...	4.797	5.727	158.7E6	749.6E6	43.341	57.355 #
9)	A Endosulfan I	5.171	6.115	148.1E6	648.5E6	43.248	47.164
10)	B gamma-Chl...	5.050	5.981	162.3E6	697.1E6	42.725	47.466
11)	B alpha-Chl...	5.114	6.059	159.5E6	684.9E6	42.916	47.234
12)	B 4,4'-DDE	5.305	6.231	151.8E6	679.8E6	44.356	47.555
13)	MA Dieldrin	5.438	6.392	158.2E6	698.3E6	43.688	47.523
14)	MA Endrin	5.715	6.624	130.7E6	509.5E6	41.287	40.732
15)	B Endosulfa...	6.011	6.850	126.4E6	568.1E6	42.067	45.820
16)	A 4,4'-DDD	5.864	6.755	129.5E6	594.2E6	45.422	49.955
17)	MA 4,4'-DDT	6.118	7.067	119.1E6	575.4E6	40.364	44.118
18)	B Endrin al...	6.192	6.982	99483824	453.5E6	45.124	46.811
19)	B Endosulfa...	6.416	7.213	126.2E6	563.0E6	43.889	47.209
20)	A Methoxychlor	6.699	7.543	64814179	314.9E6	39.293	45.680
21)	B Endrin ke...	6.927	7.706	154.7E6	655.9E6	47.504	48.861
22)	Mirex	7.115	8.168	110.7E6	427.7E6	45.196	49.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080522\
Data File : PD071370.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 16:16
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 09:36:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072622.M
Quant Title : GC Extractables
QLast Update : Tue Jul 26 03:39:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

