

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080921\
 Data File : PD064836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2021 14:50
 Operator : AR\AJ
 Sample : PB138261BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS261

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 01:02:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Jul 28 07:07:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.410	4.616	22960037	203.8E6	19.247	22.298
2)	SA Decachlor...	11.307	10.361	45794117	260.8E6	34.788	43.855 #
Target Compounds							
2)	A alpha-BHC	5.991	5.318	8142697	64426230	4.998m	5.227
3)	MA gamma-BHC...	6.389	5.738	6819202	79774835	4.348	6.996 #
4)	MA Heptachlor	7.039	6.146	9351457	68007986	5.564	5.998m
5)	MB Aldrin	7.409	6.471	7422080	59400628	4.712	5.647
6)	B beta-BHC	6.647	6.113	3713183	28580304	4.759m	4.711m
7)	B delta-BHC	6.912	6.375	6715705	51892624	3.859	4.777
8)	B Heptachlor...	7.878	7.044	7759916	51312260	4.520	5.543
9)	A Endosulfan I	8.281	7.446	7069812	42083452	5.015m	5.473
10)	B trans-Chl...	8.148	7.320	8160137	49958324	5.092	5.508
11)	B cis-Chlor...	8.228	7.390	7494102	46640913	4.810m	5.243
12)	B 4,4'-DDE	8.424	7.602	13391241	83632392	9.291	11.084
13)	MA Dieldrin	8.573	7.732	13983977	80175794	9.377	9.982
14)	MA Endrin	8.811	8.022	11936859	67230700	9.323	11.703 #
15)	B Endosulfa...	9.047	8.331	12669812	60414139	9.757	10.466
16)	A 4,4'-DDD	8.965	8.184	11882232	60760901	9.680	10.550
17)	MA 4,4'-DDT	9.283	8.445	11127136	53890976	8.524	9.790
18)	B Endrin al...	9.181	8.519	11301592	52162363	9.976m	10.569
19)	B Endosulfa...	9.421	8.748	11075010	47225228	8.036	8.761
20)	A Methoxychlor	9.773	9.037	33477512	136.1E6	46.785	53.145
21)	B Endrin ke...	9.916	9.267	13854556	67481379	9.158	11.250

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080921\
Data File : PD064836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2021 14:50
Operator : AR\AJ
Sample : PB138261BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PLCS261

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 01:02:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Jul 28 07:07:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

