

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081022\
 Data File : PD071474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2022 17:24
 Operator : AR\AJ
 Sample : PEM292
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM082

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/11/2022
 Supervised By :Ankita Jodhani 08/11/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 22:58:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081022CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 10 22:43:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.808	3.566	52554007	582.1E6	18.559	18.789
27)	SA Decachlor...	8.004	9.094	48225691	190.2E6	18.914	20.067
Target Compounds							
2)	A alpha-BHC	3.317	4.032	38866583	418.3E6	9.235	10.338
3)	MA gamma-BHC...	3.651	4.367	35337312	346.8E6	9.073	10.430
6)	B beta-BHC	3.949	4.574	16724744	138.0E6	9.835	10.027
12)	B 4,4'-DDE	5.301	6.227	1572596	8497638	0.465	0.568
14)	MA Endrin	5.707	6.617	146.0E6	584.0E6	45.371	45.350
16)	A 4,4'-DDD	5.859	6.751	3543242	16137002	1.356	1.397
17)	MA 4,4'-DDT	6.110	7.060	270.0E6	1065.8E6	96.977	92.312
18)	B Endrin al...	6.182	6.973	5046483	17238956	2.005m	1.665m
20)	A Methoxychlor	6.689	7.536	337.3E6	1208.2E6	243.806	242.455
21)	B Endrin ke...	6.917	7.699	10562869	32203756	3.062	2.569

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081022\
Data File : PD071474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2022 17:24
Operator : AR\AJ
Sample : PEM292
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 22:58:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081022CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Wed Aug 10 22:43:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By : Yogesh Patel 08/11/2022
Supervised By : Ankita Jodhani 08/11/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
PEM082