

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081022\
 Data File : PD071488.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2022 20:35
 Operator : AR\AJ
 Sample : INDB401
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB4072

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 22:29:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081022CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 10 22:16:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.807	3.564	107.7E6	1158.2E6	40.933	41.893
27)	SA Decachlor...	8.001	9.090	203.4E6	731.8E6	82.243	83.001
Target Compounds							
5)	MB Aldrin	4.280	5.285	165.4E6	807.2E6	40.526	41.791
6)	B beta-BHC	3.948	4.572	66190510	519.4E6	40.670	42.407
7)	B delta-BHC	4.182	4.820	159.4E6	1031.0E6	40.427	41.763
8)	B Heptachlo...	4.787	5.717	147.9E6	677.9E6	40.796	42.163
10)	B trans-Chl...	5.040	5.972	155.9E6	676.3E6	40.486	41.642
11)	B cis-Chlor...	5.104	6.049	152.7E6	654.9E6	40.587	41.805
12)	B 4,4'-DDE	5.296	6.222	281.0E6	1177.6E6	82.073	83.287
15)	B Endosulfa...	6.000	6.840	242.7E6	977.2E6	82.089	83.901
18)	B Endrin al...	6.181	6.971	203.0E6	811.5E6	82.508	83.873
19)	B Endosulfa...	6.406	7.204	236.8E6	915.9E6	82.272	83.218
21)	B Endrin ke...	6.915	7.696	272.8E6	981.7E6	82.325	82.964

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081022\
Data File : PD071488.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2022 20:35
Operator : AR\AJ
Sample : INDB401
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 22:29:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081022CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Wed Aug 10 22:16:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :
INDB4072