

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081324\
 Data File : PD084362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Aug 2024 15:28
 Operator : AR\AJ
 Sample : P3542-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEST-GPC2-BLANK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 01:14:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 12 16:40:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds


Target Compounds							
15) B	Endosulfa...	6.793	0.000	1924029	0	1.167	N.D. #
16) A	4,4'-DDD	6.793f	0.000	1924029	0	1.496	N.D. #
18) B	Endrin al...	6.894f	0.000	1087746	0	0.957	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081324\
Data File : PD084362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2024 15:28
Operator : AR\AJ
Sample : P3542-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 01:14:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Aug 12 16:40:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.585 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 2.930 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.793 min
Delta R.T.: -0.030 min
Response: 1924029
Conc: 1.17 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T. : 6.142 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.793 min
Delta R.T.: 0.054 min
Response: 1924029
Conc: 1.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 5.990 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 6.894 min
Delta R.T.: -0.058 min
Response: 1087746
Conc: 0.96 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T. : 6.321 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.116 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PEST-GPC2-BLANK



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.144 min
Response: 0
Conc: N.D.