

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081421\
 Data File : PD064945.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2021 06:13
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 06:32:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX081421.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 14 05:39:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	1.957	1.996 E6	-2.0	103	0.00
2	Toxaphene-1	18.357	18.330 E3	0.1	104	0.00
3	Toxaphene-2	19.804	20.320 E3	-2.6	102	0.00
4	Toxaphene-3	35.701	35.977 E3	-0.8	105	0.00
5	Toxaphene-4	34.260	36.415 E3	-6.3	104	0.00
6	Toxaphene-5	53.336	54.400 E3	-2.0	104	0.00
7 SA	Decachlorobiphenyl	2.477	2.496 E6	-0.8	103	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	20.766	21.147 E6	-1.8	103	0.00
2	Toxaphene-1	124.164	127.202 E3	-2.4	106	0.00
3	Toxaphene-2	204.829	208.827 E3	-2.0	104	0.00
4	Toxaphene-3	189.824	190.622 E3	-0.4	103	0.00
5	Toxaphene-4	344.150	348.717 E3	-1.3	103	0.00
6	Toxaphene-5	248.762	251.999 E3	-1.3	104	0.00
7 SA	Decachlorobiphenyl	11.653	11.664 E6	-0.1	103	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0