

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081421\
 Data File : PD064930.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2021 02:46
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 04:14:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081421.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 14 04:11:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.509	4.020	97645746	1010.8E6	50.000	50.000
28)	SA Decachlor...	8.280	9.106	118.0E6	537.3E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.952	4.416	153.9E6	1574.2E6	50.000	50.000
3)	MA gamma-BHC...	4.238	4.703	142.5E6	1427.5E6	50.000	50.000
4)	MA Heptachlor	4.538	5.208	149.5E6	1461.6E6	50.000	50.000
5)	MB Aldrin	4.788	5.514	147.2E6	1392.1E6	50.000	50.000
6)	B beta-BHC	4.487	4.878	61583643	625.1E6	50.000	50.000
7)	B delta-BHC	4.694	5.096	143.9E6	1414.6E6	50.000	50.000
8)	B Heptachlo...	5.233	5.904	135.6E6	1256.0E6	50.000	50.000
9)	A Endosulfan I	5.574	6.262	127.8E6	1161.2E6	50.000	50.000
10)	B gamma-Chl...	5.460	6.138	139.7E6	1281.8E6	50.000	50.000
11)	B alpha-Chl...	5.519	6.210	137.6E6	1241.4E6	50.000	50.000
12)	B 4,4'-DDE	5.683	6.368	134.0E6	1207.9E6	50.000	50.000
13)	MA Dieldrin	5.818	6.521	139.2E6	1218.3E6	50.000	50.000
14)	MA Endrin	6.077	6.741	121.4E6	962.5E6	50.000	50.000
15)	B Endosulfa...	6.350	6.953	114.4E6	942.7E6	50.000	50.000
16)	A 4,4'-DDD	6.200	6.863	116.7E6	954.4E6	50.000	50.000
17)	MA 4,4'-DDT	6.442	7.161	121.7E6	991.2E6	50.000	50.000
18)	B Endrin al...	6.520	7.078	90204413	766.4E6	50.000	50.000
19)	B Endosulfa...	6.732	7.302	114.5E6	941.9E6	50.000	50.000
20)	A Methoxychlor	6.991	7.622	68210559	463.6E6	50.000	50.000
21)	B Endrin ke...	7.228	7.776	143.1E6	934.4E6	50.000	50.000
22)	Mirex	7.415	8.226	107.2E6	605.8E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081421\
Data File : PD064930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2021 02:46
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 04:14:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081421.M
Quant Title : GC Extractables
QLast Update : Sat Aug 14 04:11:10 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

