

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081421\
 Data File : PD064931.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2021 03:00
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 14 04:22:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081421.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 14 04:11:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.509	4.020	47849987	515.9E6	24.502	25.521
28)	SA Decachlor...	8.281	9.106	61053953	284.2E6	25.860	26.446
Target Compounds							
2)	A alpha-BHC	3.952	4.416	72181942	792.2E6	23.449	25.162
3)	MA gamma-BHC...	4.238	4.703	67726729	718.6E6	23.765	25.171
4)	MA Heptachlor	4.538	5.208	70597397	743.9E6	23.605	25.450
5)	MB Aldrin	4.788	5.513	69255731	705.6E6	23.532	25.343
6)	B beta-BHC	4.487	4.879	30842372	333.7E6	25.041	26.692
7)	B delta-BHC	4.694	5.096	67097842	705.2E6	23.314	24.926
8)	B Heptachlo...	5.233	5.904	64690342	639.4E6	23.850	25.454
9)	A Endosulfan I	5.574	6.262	61055479	592.1E6	23.891	25.495
10)	B gamma-Chl...	5.460	6.138	66275919	650.1E6	23.729	25.361
11)	B alpha-Chl...	5.519	6.209	65937923	633.0E6	23.957	25.494
12)	B 4,4'-DDE	5.683	6.368	62744571	606.2E6	23.414	25.095
13)	MA Dieldrin	5.818	6.521	65338290	615.6E6	23.466	25.265
14)	MA Endrin	6.077	6.740	57723083	486.8E6	23.783	25.291
15)	B Endosulfa...	6.351	6.954	55210419	480.8E6	24.129	25.504
16)	A 4,4'-DDD	6.200	6.863	55550543	473.5E6	23.803	24.805
17)	MA 4,4'-DDT	6.442	7.162	57336459	496.2E6	23.553	25.028
18)	B Endrin al...	6.520	7.078	44601705	396.3E6	24.723	25.856
19)	B Endosulfa...	6.732	7.302	56077580	482.0E6	24.492	25.586
20)	A Methoxychlor	6.990	7.622	34929650	242.5E6	25.604	26.153
21)	B Endrin ke...	7.228	7.776	70057054	477.6E6	24.470	25.556
22)	Mirex	7.415	8.226	54577516	317.5E6	25.459	26.201

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081421\
Data File : PD064931.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2021 03:00
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 14 04:22:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081421.M
Quant Title : GC Extractables
QLast Update : Sat Aug 14 04:11:10 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

