

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081518\
 Data File : PD048935.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Aug 2018 15:10
 Operator : AJ\SJ
 Sample : J4453-13
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AB8

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 02:16:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072818CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Jul 28 05:19:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.371	3.968	22163251	317.2E6	16.086	16.492
27)	SA Decachlor...	8.067	8.993	59127474	666.1E6	36.502	37.108
Target Compounds							
3)	MA gamma-BHC...	4.089	0.000	4811180	0	2.259	N.D. #
4)	MA Heptachlor	0.000	5.096f	0	1829278	N.D.	<MDL
7)	B delta-BHC	0.000	5.056f	0	3332006	N.D.	0.131
9)	A Endosulfan I	0.000	6.126f	0	9879242	N.D.	0.437
10)	B trans-Chl...	0.000	6.126f	0	9879242	N.D.	0.397
11)	B cis-Chlor...	0.000	6.126	0	9879242	N.D.	0.412

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081518\
Data File : PD048935.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Aug 2018 15:10
Operator : AJ\SJ
Sample : J4453-13
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0AB8

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 02:16:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072818CLP.M
Quant Title : GC Extractables
QLast Update : Sat Jul 28 05:19:54 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

